

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Epstein-Barr Virus (VCA) IgG**

Prod.No.: **RE58751**

Lot: **EBVG-222**

Expiry: **2027-10-31**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0150-222	2026-04	2027-10-31
CONJ	1	KG0150-222	2026-04	2027-10-31
CONTROL -	1	NG0150-222	2026-04	2027-10-31
CUT OFF	1	CG0150-222	2026-04	2027-10-31
CONTROL +	1	PG0150-222	2026-04	2027-10-31
WASH BUF 20x	1	W-1141	2026-04	2028-04-30
SOLN STOP	1	S-1073	2026-04	2028-04-30
SUB TMB	1	T-080	2026-02	2029-02-23
DIL	1	D-501	2026-03	2028-03-31


Component	OD Value	OD Validation Criteria
CONTROL -	0.059	< 0.200 < Cut-off
CUT OFF	0.542	0.150-1.300
CONTROL +	1.110	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	19.67	10.82 - 28.52

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-05-12

Approved by: 
Quality Control

Approved on: 2026-05-12



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany