Certificate of Analysis (CoA)

Product: **Epstein-Barr Virus (VCA) IgG**
Prod.No.: **RE58751**
Lot: **EBVG-224**
Expiry: **2028-01-31**
IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0150-224	2026-07	2028-01-31
CONJ	1	KG0150-224	2026-07	2028-01-31
CONTROL -	1	NG0150-224	2026-07	2028-01-31
CUT OFF	1	CG0150-224	2026-07	2028-01-31
CONTROL +	1	PG0150-224	2026-07	2028-01-31
WASH BUF 20x	1	W-1148	2026-07	2028-07-31
SOLN STOP	1	S-1080	2026-07	2028-07-31
SUB TMB	1	T-080	2026-02	2029-02-23
DIL	1	D-503	2026-06	2028-06-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.020	< 0.200 < Cut-off
CUT OFF	0.383	0.150-1.300
CONTROL +	0.786	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.52	12.39 - 32.65

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-08-25

Approved by: _____
Quality Control

Approved on: 2026-08-25

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany