Certificate of Analysis (CoA)

Product: **Epstein-Barr Virus (VCA) IgM**
Prod.No.: **RE58761**
Lot: **EBVM-226**
Expiry: **2028-02-29**
IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0150-226	2026-08	2028-02-29
CONJ	1	KM0150-226	2026-08	2028-02-29
CONTROL -	1	NM0150-226	2026-08	2028-02-29
CUT OFF	1	CM0150-226	2026-08	2028-02-29
CONTROL +	1	PM0150-226	2026-08	2028-02-29
WASH BUF 20x	1	W-1151	2026-09	2028-09-30
SOLN STOP	1	S-1083	2026-09	2028-09-30
SUB TMB	1	T-080	2026-02	2029-02-23
DIL M	1	DM-356	2026-07	2028-07-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.010	< 0.200 < Cut-off
CUT OFF	0.498	0.150-1.300
CONTROL +	1.176	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	24.73	13.60 - 35.86

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-10-06

Approved by: _____

Quality Control

Approved on: 2026-10-07

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany