

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Mycoplasma pneumoniae IgG**

Prod.No.: **RE58771**

Lot: **MYCG-197N**

Expiry: **2026-07-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0350-197	2025-01	2026-07-31
CONJ	1	KG0350-197	2025-01	2026-07-31
CONTROL -	1	NG0350-197	2025-01	2026-07-31
CUT OFF	1	CG0350-197	2025-01	2026-07-31
CONTROL +	1	PG0350-197	2025-01	2026-07-31
WASH BUF 20x	1	W-1103	2024-10	2026-10-31
SOLN STOP	1	S-1034	2024-10	2026-10-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-489	2024-10	2026-10-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.115	< 0.200 < Cut-off
CUT OFF	0.525	0.150-1.300
CONTROL +	1.166	> Cut-off

Final Report issued: 2025-01-31

Approved by: 
Quality Control

Approved on: 2025-01-31



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany