Certificate of Analysis (CoA)

Product: **Mycoplasma pneumoniae IgA**
Prod.No.: **RE58791**
Lot: **MYCA-205**
Expiry: **2027-07-31**
IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MA0350-205	2026-01	2027-07-31
CONJ	1	KA0350-205	2026-01	2027-07-31
CONTROL -	1	NA0350-205	2026-01	2027-07-31
CUT OFF	1	CA0350-205	2026-01	2027-07-31
CONTROL +	1	PA0350-205	2026-01	2027-07-31
WASH BUF 20x	1	W-1135	2026-02	2028-02-29
SOLN STOP	1	S-1067	2026-01	2028-01-31
SUB TMB	1	T-079	2025-12	2028-12-10
DIL	1	D-499	2026-01	2028-01-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.010	< 0.200 < Cut-off
CUT OFF	0.589	0.150-1.300
CONTROL +	1.107	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	16.71	>10 - 24.23

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-02-26

Approved by: Quality Control

Approved on: 2026-02-26

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany