

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Chikungunya Virus IgM μ -capture**

Prod.No.: **RE58841**

Lot: **CHIM-116N**

Expiry: **2026-05-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0590-116	2024-11	2026-05-31
CONJ STRE	1	KM0590-116	2024-11	2026-05-31
AG LYO	6	AGM0590-116	2024-11	2026-05-31
AB SOLN	1	ABM0590-116	2024-11	2026-05-31
CONTROL -	1	NM0590-116	2024-11	2026-05-31
CUT OFF	1	CM0590-116	2024-11	2026-05-31
CONTROL +	1	PM0590-116	2024-11	2026-05-31
WASH BUF 20x	1	W-1101	2024-10	2026-10-31
SOLN STOP	1	S-1032	2024-09	2026-09-30
SUB TMB	1	T-076	2022-05	2026-11-30
DIL	1	D-488	2024-08	2026-08-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.238	< Cut-off
CUT OFF	0.618	0.150-1.300
CONTROL +	1.857	> Cut-off

Final Report issued: 2024-12-17

Approved by: _____

Quality Control



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany

Approved on: 2024-12-17



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany