

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Coxiella burnetii (Q-Fever) Phase 1 IgG**

Prod.No.: **RE58851**

Lot: **COX1G-053N**

Expiry: **2026-09-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	M1G0600-053	2025-03	2026-09-30
CONJ	1	K1G0600-053	2025-03	2026-09-30
CONTROL -	1	N1G0600-053	2025-03	2026-09-30
CUT OFF	1	C1G0600-053	2025-03	2026-09-30
CONTROL +	1	P1G0600-053	2025-03	2026-09-30
WASH BUF 20x	1	W-1108	2025-02	2027-02-28
SOLN STOP	1	S-1039	2025-01	2027-01-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-490	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.006	< 0.200 < Cut-off
CUT OFF	0.598	0.150-1.300
CONTROL +	1.379	> Cut-off

Final Report issued: 2025-04-04

Approved by: 
Quality Control

Approved on: 2025-04-04



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany