

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Leptospira IgG**
 Prod.No.: **RE58931**
 Lot: **LEPG-092**
 Expiry: **2027-06-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0660-092	2025-12	2027-06-30
CONJ	1	KG0660-092	2025-12	2027-06-30
CONTROL -	1	NG0660-092	2025-12	2027-06-30
CUT OFF	1	CG0660-092	2025-12	2027-06-30
CONTROL +	1	PG0660-092	2025-12	2027-06-30
WASH BUF 20x	1	W-1129	2025-11	2027-11-30
SOLN STOP	1	S-1064	2025-11	2027-11-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-496	2025-09	2027-09-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.035	< 0.200 < Cut-off
CUT OFF	0.694	0.150-1.300
CONTROL +	1.331	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	17.83	10.0 - 25.85

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-01-06

Approved by: 
 Quality Control

Approved on: 2026-01-06



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany