

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Leptospira IgM**

Prod.No.: **RE58941**

Lot: **LEPM-088N**

Expiry: **2026-09-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0660-088	2025-03	2026-09-30
CONJ	1	KM0660-088	2025-03	2026-09-30
CONTROL -	1	NM0660-088	2025-03	2026-09-30
CUT OFF	1	CM0660-088	2025-03	2026-09-30
CONTROL +	1	PM0660-088	2025-03	2026-09-30
WASH BUF 20x	1	W-1106	2025-01	2027-01-31
SOLN STOP	1	S-1037	2024-11	2026-11-30
SUB TMB	1	T-077	2022-09	2027-03-31
DIL M	1	DM-342	2024-11	2026-11-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.018	< 0.200 < Cut-off
CUT OFF	0.583	0.150-1.300
CONTROL +	0.950	> Cut-off

Final Report issued: 2025-03-25

Approved by: *S. Herrnschütz*
Quality Control

Approved on: 2025-03-25



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany