Certificate of Analysis (CoA)

Product: **Leptospira IgM**
Prod.No.: **RE58941**
Lot: **LEPM-098**
Expiry: **2028-02-29**
IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0660-098	2026-08	2028-02-29
CONJ	1	KM0660-098	2026-08	2028-02-29
CONTROL -	1	NM0660-098	2026-08	2028-02-29
CUT OFF	1	CM0660-098	2026-08	2028-02-29
CONTROL +	1	PM0660-098	2026-08	2028-02-29
WASH BUF 20x	1	W-1148	2026-07	2028-07-31
SOLN STOP	1	S-1080	2026-07	2028-07-31
SUB TMB	1	T-080	2026-02	2029-02-23
DIL M	1	DM-356	2026-07	2028-07-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.029	< 0.200 < Cut-off
CUT OFF	0.607	0.150-1.300
CONTROL +	0.988	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	16.59	>10 - 24.06

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-09-04

Approved by: _____
Quality Control

Approved on: 2026-09-04

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany