

# Certificate of Analysis (CoA)



**GOLD  
STANDARD  
DIAGNOSTICS**



Product: **Strongyloides**

Prod.No.: **RE58951**

Lot: **STRO-053**

Expiry: **2026-11-30**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
<b>MTP</b>	1	MAB0690-053	2025-05	2026-11-30
<b>CONJL</b>	1	KAB0690-053	2025-05	2026-11-30
<b>CONTROL -</b>	1	NAB0690-053	2025-05	2026-11-30
<b>CUT OFF</b>	1	CAB0690-053	2025-05	2026-11-30
<b>CONTROL +</b>	1	PAB0690-053	2025-05	2026-11-30
<b>WASH BUF 20x</b>	1	W-1111	2025-03	2027-03-31
<b>SOLN STOP</b>	1	S-1042	2025-02	2027-02-28
<b>SUB TMB</b>	1	T-077	2022-09	2027-03-31
<b>DIL</b>	1	D-491	2025-02	2027-02-28

Component	OD Value	OD Validation Criteria
<b>CONTROL -</b>	0.032	< 0.200 < Cut-off
<b>CUT OFF</b>	0.530	0.150-1.300
<b>CONTROL +</b>	1.126	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
<b>CONTROL +</b>	20.84	11.46 - 30.21

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-06-06

Approved by: ACc  
Quality Control

Approved on: 2025-06-06



Gold Standard Diagnostics Frankfurt GmbH  
Waldstrasse 23 A6  
63128 Dietzenbach, Germany



IBL International GmbH  
Flughafenstrasse 52a, 22335 Hamburg  
Germany