

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Strongyloides**
 Prod.No.: **RE58951**
 Lot: **STRO-055**
 Expiry: **2027-03-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MAB0690-055	2025-09	2027-03-31
CONJL	1	KAB0690-055	2025-09	2027-03-31
CONTROL -	1	NAB0690-055	2025-09	2027-03-31
CUT OFF	1	CAB0690-055	2025-09	2027-03-31
CONTROL +	1	PAB0690-055	2025-09	2027-03-31
WASH BUF 20x	1	W-1120	2025-08	2027-08-31
SOLN STOP	1	S-1051	2025-08	2027-08-31
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-494	2025-06	2027-06-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.011	< 0.200 < Cut-off
CUT OFF	0.583	0.150-1.300
CONTROL +	1.309	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	23.78	13.08 - 34.48

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-10-16

Approved by: 
 Quality Control

Approved on: 2025-10-16



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany