

Quality Control Certificate




Serotonin ELISA

REF RE59121	LOT ESE162	2026-09-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration ng/mL	Measured value OD	manual	
			Range OD	OD/ODmax
CAL A	0	2.452	> 1.000	100%
CAL B	0.08	2.133	-	87%
CAL C	0.24	1.656	-	68%
CAL D	0.73	1.072	-	44%
CAL E	2.2	0.536	-	22%
CAL F	6.6	0.256	-	10%
CAL G	19.8	0.125	< 0.500	5%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	76	46	77	107
CONTROL 2	272	190	271	353



Kit Composition:

REF		LOT	
KESE651	LYO CONTROL 1	K419	2027-10
KESE661	LYO CONTROL 2	K420	2027-10
KEZZ811	MTP	K346	2028-09
KESE521	ANTISERUM	L955-A	2026-09
KESE531	BIOTIN	L954-A	2026-10
KESE541	CONC ENZCONJ	K390	2027-09
KESE611	CAL A-G	K383	2026-09

REF		LOT	
KESE761	ACYLREAG	K378	2026-09
KEZZ681	CONC ASSAYBUF	L957-A	2027-09
KEZZ701	CONC WASHBUF	K377	2027-09
KEZZ961	PNPP SUBS	K180-2	2027-09
KEZZ951	PNPP STOP	K381	2028-09

Valid edition of the instructions for use: V2025-01

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T.Schmidt

Accepted

Validation date: 2025-11-21

This document has been produced electronically and is valid without a signature.