

# Quality Control Certificate




## 5-HIAA (5-Hydroxy-3-indole Acetic Acid) ELISA

<b>REF</b> RE59131	<b>LOT</b> EIN116	<b>2026-10-31</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

### Final Release Results:

	Concentration mg/L	Measured value OD	Range OD	OD/ODmax
CAL A	0.00	2.848	> 1.300	100%
CAL B	0.2	2.656	-	93%
CAL C	0.55	2.427	-	85%
CAL D	1.8	2.023	-	71%
CAL E	7.5	1.530	-	54%
CAL F	20	1.068	-	38%
CAL G	55	0.674	-	24%

	Concentration found mg/L	Acceptable range		
		lower limit mg/L	target mg/L	upper limit mg/L
CONTROL 1	6.3	4.3	6.2	8.0
CONTROL 2	33.8	24.0	33.3	42.6



### Kit Composition:

REF		LOT	
KEIN651	CONTROL 1	K220	2026-12
KEIN661	CONTROL 2	K221	2026-12
KEZZ811	MTP	L436	2027-07
KEIN521	ANTISERUM	K218	2026-10
KEIN531	BIOTIN	L501-A	2026-10
KEIN541	CONC ENZCONJ	K226	2027-06
KEIN611	CAL A-G	M525B-3B	2026-10

REF		LOT	
KEZZ681	CONC ASSAYBUF	L950-A	2027-06
KEIN771	DILREAG	K217	2026-10
KEIN751	METHYLREAG	K219	2026-10
KEZZ701	CONC WASHBUF	L939-2	2027-03
KEZZ961	PNPP SUBS	K180	2027-09
KEZZ951	PNPP STOP	L193	2027-02
KEIN761	HCl	K190	2028-05

Valid edition of the instructions for use: V2014-11

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2025-09-30

**Accepted**

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)