

# Quality Control Certificate




## Dopamine ELISA

<b>REF</b> RE59161	<b>LOT</b> EDO156	<b>2026-06-30</b> YYYY-MM-DD
--------------------	-------------------	---------------------------------

### Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.096	-	5%
CAL B	60	0.168	-	9%
CAL C	180	0.284	-	14%
CAL D	585	0.678	-	34%
CAL E	2300	1.444	-	73%
CAL F	11470	1.973	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	157	98	163	228
CONTROL 2	649	360	600	840



### Kit Composition:

REF		LOT	⌚
KEZZ521	CONTROL 1	K212	2027-06
KEZZ531	CONTROL 2	K213	2027-06
KEZZ801	MTP	K209	2028-06
KEZZ431	ANTISERUM DO	K207	2026-07
KEZZ821	CONC ENZCONJ	K214	2027-07
KEZZ511	CAL A-F	K211	2027-06
KEZZ541	ACYLREAG	K208	2026-07
KEZZ555	EXTRPLATE	K210	2027-06
KEZZ565	EXTRBUF	K199	2026-07

REF		LOT	⌚
KEZZ575	LYO COMT	K204	2026-06
KEZZ581	COENZ	K203	2026-06
KEZZ591	ENZBUF	K201	2028-06
KEZZ422	COMT ADD	K202	2026-07
KEZZ609	RELEASEBUF	K200	2026-07
KLZZ879	CONC WASHBUF	K215	2027-06
KEZZ971	PNPP SUBS	L747-4	2026-08
KEZZ951	PNPP STOP	K216	2028-06

Valid edition of the instructions for use: 2014-12

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T.Schmidt

Validation date: 2025-09-01

**Accepted**

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)