

Quality Control Certificate




Normetanephine ELISA

REF RE59171	LOT ENM201	2026-06-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration µg/L	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.918	> 1.000	100%
CAL B	77	2.404	-	82%
CAL C	192	1.916	-	66%
CAL D	480	1.484	-	51%
CAL E	1200	0.913	-	31%
CAL F	3000	0.547	-	19%
CAL G	7500	0.352	< 0.500	12%



	Concentration found µg/L	Acceptable range		
		lower limit µg/L	target µg/L	upper limit µg/L
CONTROL 1	214	125	195	266
CONTROL 2	685	448	700	951

Kit Composition:

REF		LOT	⌚
KEZZ621	CONTROL 1	L789	2026-12
KEZZ631	CONTROL 2	L790	2026-12
KEZZ811	MTP	L337	2027-04
KEZZ661	HYDRTUB	L748	2029-11
KEZZ671	HCl	L472	2027-07
KEZZ651	INDICATORBUF	L792	2027-12
KEZZ641	ACYLREAG	L791	2026-12
KENM541	CONC ENZCONJ	L787	2027-01

REF		LOT	⌚
KEZZ611	CAL A-G	L788	2026-12
KENM533	BIOTIN	L360-A	2026-06
KENM521	ANTISERUM	L786	2026-06
KEZZ681	CONC ASSAYBUF	L455-B	2026-11
KLZZ879	CONC WASHBUF	L738	2026-11
KEZZ961	PNPP SUBS	L747-2	2026-08
KEZZ951	PNPP STOP	L737	2027-11

Valid edition of the instructions for use: 2024-04

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Validation date: 2025-03-03

Accepted

This document has been produced electronically and is valid without a signature.