

# Quality Control Certificate



## MetCombi Plasma ELISA (Metanephrine)


<b>REF</b> RE59202	<b>LOT</b> MNPE150	<b>2026-11-30</b> YYYY-MM-DD
--------------------	--------------------	---------------------------------

### Final Release Results:

	Concentration pg/mL	manual			automated e.g. DSX
		Measured value OD	Range OD	OD/ODmax	Measured value OD
CAL A	0	1.944	> 1.000	100%	2.389
CAL B	18	1.407	-	72%	1.791
CAL C	45	1.186	-	61%	1.404
CAL D	120	0.947	-	49%	1.147
CAL E	340	0.719	-	37%	0.932
CAL F	700	0.590	-	30%	0.763

	Acceptable range			Concentration found	
	lower limit pg/mL	target pg/mL	upper limit pg/mL	pg/mL	
				manual	automated
CONTROL 1	44	74	104	72	58
CONTROL 2	132	220	308	216	150

\*During automated procedure the optical densities of calibrators and controls may rise due to temperature and specific device conditions. In case of none conformance with the OD ranges stated on the QC Certificate please contact IBL-International or use the validation criteria only!



### Kit Composition:

REF	LOT		REF	LOT			
KEWP501	MTP MN	MPMT48	2027-03	KEWP521	ANTISERUM MN	MPAS50	2027-02
KEWP611A	LYO CAL A	NPST46	2027-03	KEWP541	ENZCONJ	NPKJ48	2027-03
KEWP611	LYO CAL B-F	NPST46	2027-03	KEWP551	SOLVENT	NPOS48	2027-03
KEWP651	LYO CONTROL 1	NPKO46	2027-03	KEWP761	ACYLBUF	NPAP41	2029-03
KEWP661	LYO CONTROL 2	NPKO46	2027-03	KEWP791	PREC TUBES	NPPR44	2026-11
KEWP771	PREC REAG 1	NPP142	2029-07	KEWP561	WASHBUF CONC	NPWP32	2028-04
KEWP781	PREC REAG 2	NPP245	2029-07	KEWP571	TMB SUBS	NPSU49	2027-03
KEWP751	LYO ACYLREAG	NPAR48	2027-03	KEWP581	TMB STOP	NPSO43	2029-03

Valid edition of the instructions for use: Ver.9 / 12-2024

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

# Accepted

Validation date: 2025-06-12

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)