

Quality Control Certificate




MetCombi Plasma ELISA (Normetanephrine)

REF RE59202	LOT MNPE150	2026-11-30 YYYY-MM-DD
--------------------	--------------------	---------------------------------

Final Release Results:

	Concentration pg/mL	manual			e.g. automated/ DSX
		Measured value OD	Range OD	OD/ODmax	Measured value OD
CAL A	0	1.815	> 1.000	100%	2.336
CAL B	30	1.544	-	85%	1.726
CAL C	65	1.249	-	69%	1.417
CAL D	120	0.986	-	54%	1.023
CAL E	240	0.707	-	39%	0.818
CAL F	500	0.532	-	29%	0.584



	Acceptable range			Concentration found	
	lower limit pg/mL	target pg/mL	upper limit pg/mL	pg/mL	
				manual	automated
CONTROL 1	58	97	136	99	90
CONTROL 2	108	180	252	198	159

*During automated procedure the optical densities of calibrators and controls may rise due to temperature and specific device conditions. In case of none conformance with the OD ranges stated on the QC Certificate please contact IBL-International or use the validation criteria only!

Kit Composition:

REF	LOT		REF	LOT	
KEWP511	MTP NMN	NPMT48 2027-03	KEWP531	ANTISERUM NMN	NPAS50 2027-02
KEWP611A	LYO CAL A	NPST46 2027-03	KEWP541	ENZCONJ	NPKJ48 2027-03
KEWP611	LYO CAL B-F	NPST46 2027-03	KEWP551	SOLVENT	NPOS48 2027-03
KEWP651	LYO CONTROL 1	NPKO46 2027-03	KEWP761	ACYLBUF	NPAP41 2029-03
KEWP661	LYO CONTROL 2	NPKO46 2027-03	KEWP791	PREC TUBES	NPPR44 2026-11
KEWP771	PREC REAG 1	NPP142 2029-07	KEWP561	WASHBUF CONC	NPWP32 2028-04
KEWP781	PREC REAG 2	NPP245 2029-07	KEWP571	TMB SUBS	NPSU49 2027-03
KEWP751	LYO ACYLREAG	NPAR48 2027-03	KEWP581	TMB STOP	NPSO43 2029-03

Valid edition of the instructions for use: Ver.9 / 12-2024

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Accepted

Validation date: 2025-06-12

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl