

Quality Control Certificate




MetCombi Plasma ELISA (Metanephrine)

REF	RE59202	LOT	MNPE152		2027-12-31 YYYY-MM-DD
------------	----------------	------------	----------------	--	---------------------------------

Final Release Results:

	Concentration pg/mL	manual			automated e.g. DSX
		Measured value OD	Range OD	OD/ODmax	Measured value OD
CAL A	0	1.724	> 1.000	100%	2.633
CAL B	18	1.279	-	74%	1.975
CAL C	48	1.106	-	64%	1.605
CAL D	130	0.841	-	49%	1.263
CAL E	310	0.711	-	41%	1.069
CAL F	750	0.590	-	34%	0.860



	Acceptable range			Concentration found	
	lower limit pg/mL	target pg/mL	upper limit pg/mL	pg/mL	
				manual	automated
CONTROL 1	47	78	109	79	68
CONTROL 2	107	178	249	212	189

*During automated procedure the optical densities of calibrators and controls may rise due to temperature and specific device conditions. In case of none conformance with the OD ranges stated on the QC Certificate please contact IBL-International or use the validation criteria only!

Kit Composition:

REF	LOT		REF	LOT			
KEWP501	MTP MN	MPMT50	2028-01	KEWP521	ANTISERUM MN	MPAS52	2027-12
KEWP611A	LYO CAL A	NPST48	2028-01	KEWP541	ENZCONJ	NPKJ50	2028-02
KEWP611	LYO CAL B-F	NPST48	2028-01	KEWP551	SOLVENT	NPOS50	2028-01
KEWP651	LYO CONTROL 1	NPKO48	2028-01	KEWP761	ACYLBUF	NPAP43	2030-02
KEWP661	LYO CONTROL 2	NPKO48	2028-01	KEWP791	PREC TUBES	NPPR46	2028-06
KEWP771	PREC REAG 1	NPP144	2030-12	KEWP561	WASHBUF CONC	NPWP34	2029-01
KEWP781	PREC REAG 2	NPP247	2031-02	KEWP571	TMB SUBS	NPSU51	2027-12
KEWP751	LYO ACYLREAG	NPAR50	2028-01	KEWP581	TMB STOP	NPSO45	2028-08

Valid edition of the instructions for use: Ver.9 / 12-2024

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Accepted

Validation date: 2026-04-10

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl