

Quality Control Certificate

Adrenalin (Epinephrine) ELISA




REF RE59251	LOT EAD185	2026-06-30 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.133	-	6%
CAL B	1.5	0.164	-	8%
CAL C	5.0	0.265	-	13%
CAL D	15	0.536	-	25%
CAL E	50	1.127	-	54%
CAL F	150	2.103	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	6.6	3.7	6.1	8.6
CONTROL 2	32.3	19.5	28.7	37.9



Kit Composition:

REF	LOT	2026-06	REF	LOT	2026-06		
KEZZ521	CONTROL 1	K212	2027-06	KEZZ575	LYO COMT	K204	2026-06
KEZZ531	CONTROL 2	K213	2027-06	KEZZ581	COENZ	K203	2026-06
KEZZ801	MTP	K209	2028-06	KEZZ591	ENZBUF	K201	2028-06
KEZZ401	ANTISERUM AD	K205	2026-07	KEZZ422	COMT ADD	K202	2026-07
KEZZ821	CONC ENZCONJ	K214	2027-07	KEZZ601/9	RELEASEBUF	K200	2026-07
KEZZ511	CAL A-F	K211	2027-06	KLZZ879	CONC WASHBUF	K215	2027-06
KEZZ541	ACYLREAG	K208	2026-07	KEZZ971	PNPP SUBS	L747-4	2026-08
KEZZ555	EXTRPLATE	K210	2027-06	KEZZ951	PNPP STOP	K216	2028-06
KEZZ565	EXTRBUF	K199	2026-07				

Valid edition of the instructions for use: 2014-12

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T.Schmidt

Accepted

Validation date: 2025-09-01

This document has been produced electronically and is valid without a signature.