

Quality Control Certificate

Adrenalin (Epinephrine) ELISA



REF RE59251


LOT EAD186

2027-02-28
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.086	-	4%
CAL B	1.5	0.149	-	7%
CAL C	5.0	0.261	-	13%
CAL D	15	0.561	-	27%
CAL E	50	1.173	-	56%
CAL F	150	2.088	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	7.5	4.1	6.9	9.7
CONTROL 2	32.9	21.4	31.5	41.6



Kit Composition:

REF	LOT		REF	LOT			
KEZZ521	CONTROL 1	K627	2028-02	KEZZ575	LYO COMT	K622	2027-02
KEZZ531	CONTROL 2	K628	2028-02	KEZZ581	COENZ	K621	2027-02
KEZZ801	MTP	K614	2029-02	KEZZ591	ENZBUF	K619	2029-02
KEZZ401	ANTISERUM AD	K618	2027-03	KEZZ422	COMT ADD	K620	2027-03
KEZZ821	CONC ENZCONJ	K629	2028-03	KEZZ601/9	RELEASEBUF	K624	2027-03
KEZZ511	CAL A-F	K626	2028-02	KLZZ879	CONC WASHBUF	K586	2028-01
KEZZ541	ACYLREAG	K623	2027-03	KEZZ971	PNPP SUBS	K492-3	2027-11
KEZZ555	EXTRPLATE	K615	2028-02	KEZZ951	PNPP STOP	K630	2029-03
KEZZ565	EXTRBUF	K625	2027-03				

Valid edition of the instructions for use: 2014-12

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-04-24

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl