

Quality Control Certificate




Neopterin ELISA

REF RE59321**LOT** ENO243**2028-07-31**
YYYY-MM-DD

Final Release Results:

	Concentration nmol/L	Measured value OD	Range OD	OD/ODmax
CAL A	0	1.919	> 1.000	100%
CAL B	1.35	1.624	-	85%
CAL C	4	1.240	-	65%
CAL D	12	0.793	-	41%
CAL E	37	0.496	-	26%
CAL F	111	0.317	-	17%

	Concentration found nmol/L	Acceptable range lower limit nmol/L	target nmol/L	upper limit nmol/L
CONTROL 1	6.4	3.7	6.1	8.5
CONTROL 2	20.2	13.6	21.0	28.3



Kit Composition:

REF		LOT	
KENO651	CONTROL 1	K900	2028-07
KENO661	CONTROL 2	K901	2028-07
KEZZ811	MTP	K896	2029-07
KENO521	ANTISERUM	K897	2028-07
KENO541	ENZCONJ	K898	2028-07
KENO611	CAL A-F	K899	2028-07
KENO751	ASSAYBUF	K902	2028-08
KEZZ751	CONC WASHBUF	K889	2028-08

REF		LOT	
KEZZ981	TMB SUBS	K610-4	2029-08
KENO781	TMB STOP	K903	2029-08

Valid edition of the instructions for use: 2024-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-09-28

Accepted

This document has been produced electronically and is valid without a signature.

**IBL International GmbH**
Flughafenstr. 52A, 22335 Hamburg, GermanyTel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl