

Quality Control Certificate




Neopterin ELISA

REF RE59355D	LOT ENED136	2026-10-31 YYYY-MM-DD
---------------------	--------------------	---------------------------------

Final Release Results:

	Concentration nmol/L	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.332	-	100%
CAL B	1.35	1.876	-	80%
CAL C	4	1.312	-	56%
CAL D	12	0.891	-	38%
CAL E	37	0.589	< 1.200	25%
CAL F	111	0.378	-	16%

	Concentration found nmol/L	Acceptable range		
		lower limit nmol/L	target nmol/L	upper limit nmol/L
CONTROL 1	4.9	2.4	4.9	7.4
CONTROL 2	17.3	11.1	17.0	23.0



Kit Composition:

REF		LOT	
KENB555	CONC ENZCONJ	L703	2026-10
KENO611	CAL A-F	L699	2026-11
KENB511	MTP	L698	2026-10
KENB855	ASSAYBUF	L702	2026-10
KENB765	CONC WASHBUF	L706	2026-12
KEZZ991	TMB SUBS	L704	2027-10
KEZZ921	TMB STOP	L705	2027-11

Control Set:

REF		LOT	
KENO651	CONTROL 1	L700	2026-11
KENO661	CONTROL 2	L701	2026-11
KENB851	ENZCONJDIL	L702	2026-10

Valid edition of the instructions for use: V2023-04

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Validation date: 2025-01-21

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl