

# Quality Control Certificate



## Neopterin ELISA

REF RE59355D


LOT ENED138

2027-01-31  
YYYY-MM-DD

### Final Release Results:

	Concentration nmol/L	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.740	-	100%
CAL B	1.35	2.196	-	80%
CAL C	4	1.519	-	55%
CAL D	12	1.000	-	36%
CAL E	37	0.639	< 1.200	23%
CAL F	111	0.416	-	15%

	Concentration found nmol/L	Acceptable range		
		lower limit nmol/L	target nmol/L	upper limit nmol/L
CONTROL 1	5.4	2.5	5.0	7.5
CONTROL 2	16.8	11.7	18.1	24.4



### Kit Composition:

REF		LOT	
KENB555	CONC ENZCONJ	L897	2027-01
KENO611	CAL A-F	L893	2027-02
KENB511	MTP	L846	2027-01
KENB855	ASSAYBUF	L896	2027-01
KENB765	CONC WASHBUF	L854	2027-02
KEZZ991	TMB SUBS	L541-5	2027-10
KEZZ921	TMB STOP	L853	2028-01

### Control Set:

REF		LOT	
KENO651	CONTROL 1	L894	2027-02
KENO661	CONTROL 2	L895	2027-02
KENB851	ENZCONJDIL	L896	2027-01

Valid edition of the instructions for use: V2023-04

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-04-11

# Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)