

# Quality Control Certificate



## Neopterin ELISA

REF RE59355D


LOT ENED139

2027-02-28  
YYYY-MM-DD

### Final Release Results:

	Concentration nmol/L	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.405	-	100%
CAL B	1.35	2.065	-	86%
CAL C	4	1.349	-	56%
CAL D	12	0.939	-	39%
CAL E	37	0.593	< 1.200	25%
CAL F	111	0.375	-	16%

	Concentration found nmol/L	Acceptable range lower limit nmol/L	target nmol/L	upper limit nmol/L
CONTROL 1	4.9	2.3	4.8	7.3
CONTROL 2	17.3	11.4	17.5	23.6



### Kit Composition:

REF		LOT	
KENB555	CONC ENZCONJ	L925	2027-02
KENO611	CAL A-F	L921	2027-03
KENB511	MTP	L847	2028-02
KENB855	ASSAYBUF	L924	2027-02
KENB765	CONC WASHBUF	L854	2027-02
KEZZ991	TMB SUBS	L541-5	2027-10
KEZZ921	TMB STOP	L853	2028-01

### Control Set:

REF		LOT	
KENO651	CONTROL 1	L922	2027-03
KENO661	CONTROL 2	L923	2027-03
KENB851	ENZCONJDIL	L924	2027-02

Valid edition of the instructions for use: V2023-04

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-05-13

# Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)