

# Quality Control Certificate




## Testosterone Luminescence Immunoassay

<b>REF</b> RE62031 & RE62039.	<b>LOT</b> LTE207	<b>2027-01-31</b> YYYY-MM-DD
-------------------------------	-------------------	---------------------------------

**Final Release Results:**

	Concentration pg/mL	Measured value RLU	Range RLU	RLU/RLUmax
CAL A	0	2544913	-	100%
CAL B	6.4	2087123	-	82%
CAL C	16	1807883	-	71%
CAL D	40	1203448	-	47%
CAL E	100	689638	-	27%
CAL F	250	355070	-	14%
CAL G	760	134445	-	5%



	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL 1	32.5	19.4	29.9	40.4
CONTROL 2	201	162	203	243

**Kit Composition:**

REF		LOT	
KLTE651/9	CONTROL 1	K177	2027-01
KLTE661/9	CONTROL 2	K178	2027-01
KLZZ511	MTP	K123	2027-10
KLTE541/9	CONC ENZCONJ	K238	2027-01
KLTE611/9	CAL A	K176	2027-01
KLTE621/9	CAL B-G	K176	2027-01

REF		LOT	
KLTE521/9	ANTISERUM	K175	2027-01
KLZZ681/9	ASSAYBUF	K156	2027-05
KLZZ601/9	LUMINREAG AP	K241	2027-04
KLZZ879	CONC WASHBUF	L951	2027-04

Valid edition of the instructions for use: 2023-12

**This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.**

Verified by: **Dr. Spangenberg**

Validation date: **2025-08-05**

Accepted

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)