

Quality Control Certificate




Cortisol Saliva Luminescence Immunoassay

REF RE62111 & RE62119	LOT LCL210	2027-11-30 YYYY-MM-DD
------------------------------	-------------------	---------------------------------

Final Release Results:

	Concentration µg/dL	Measured value RLU	Range RLU	RLU/RLUmax
CAL A	0	27420685	-	100%
CAL B	0.015	20392473	-	74%
CAL C	0.03	16604290	-	61%
CAL D	0.10	8775608	-	32%
CAL E	0.40	2582170	-	9%
CAL F	1.00	1004870	-	4%
CAL G	3.20	240440	-	1%

	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.078	0.052	0.079	0.107
CONTROL 2	0.74	0.50	0.72	0.94



Kit Composition:

REF		LOT	
KLCL651/9	CONTROL 1	K657	2028-04
KLCL661/9	CONTROL 2	K658	2028-04
KLCL511	MTP	K680	2028-04
KLCL541/9	ENZCONJ	K720	2028-05
KLCL611/9	CAL A	K655	2028-04
KLCL621/9	CAL B-G	K656	2028-04
KLZZ709	CONC WASHBUF	K602	2028-03

REF		LOT	
KLZZ641/9	LUMINREAG 1	K721	2028-01
KLZZ651/9	LUMINREAG 2	K722	2028-01

Valid edition of the instructions for use: 2022-10

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-08

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl