

Quality Control Certificate



Cortisol Saliva Luminescence Immunoassay

REF RE62111 & RE62119


LOT LCL211

2028-02-29
YYYY-MM-DD

Final Release Results:

	Concentration µg/dL	Measured value RLU	Range RLU	RLU/RLUmax
CAL A	0	23146753	-	100%
CAL B	0.015	17210990	-	74%
CAL C	0.03	13522990	-	58%
CAL D	0.10	7026388	-	30%
CAL E	0.40	2288778	-	10%
CAL F	1.00	819503	-	4%
CAL G	3.20	214303	-	1%

	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.086	0.059	0.090	0.122
CONTROL 2	0.79	0.52	0.75	0.97



Kit Composition:

REF		LOT	
KLCL651/9	CONTROL 1	K284A-2A	2028-08
KLCL661/9	CONTROL 2	K285A-2A	2028-08
KLCL511	MTP	K819	2028-06
KLCL541/9	ENZCONJ	K866	2028-08
KLCL611/9	CAL A	K905	2028-08
KLCL621/9	CAL B-G	K283A-2A	2028-08
KLZZ709	CONC WASHBUF	K867	2028-08

REF		LOT	
KLZZ641/9	LUMINREAG 1	K868	2028-05
KLZZ651/9	LUMINREAG 2	K869	2028-05

Valid edition of the instructions for use: 2022-10

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-08-03

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl