

Quality Control Certificate




17beta-Estradiol Luminescence Immunoassay

REF RE62149	LOT LEL138	2026-07-31 YYYY-MM-DD
--------------------	-------------------	---------------------------------

Final Release Results:

	Concentration pg/mL	Measured value RLU	Range RLU	RLU/RLUmax
CAL A	0	2081510	-	100%
CAL B	0.9	2021500	-	97%
CAL C	2	1782690	-	86%
CAL D	4	1573243	-	76%
CAL E	8	1120475	-	54%
CAL F	16	502593	-	24%
CAL G	64	239455	-	12%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL 1	3.3	2.1	3.2	4.4
CONTROL 2	13.1	10.4	13.0	15.6



Kit Composition:

REF		LOT	
KLEL659	CONTROL 1	L734	2026-07
KLEL669	CONTROL 2	L735	2026-07
KLEL511	MTP	L683	2027-04
KLEL549	CONC ENZCONJ	L812	2026-07
KLEL619	CAL A	L733	2026-07
KLEL629	CAL B-G	L733	2026-07

REF		LOT	
KLEL529	ANTISERUM	L784	2026-07
KLZZ689	ASSAYBUF	L721	2026-11
KLZZ609	LUMINREAG AP	L813	2026-12
KLZZ879	CONC WASHBUF	L727	2026-11

Valid edition of the instructions for use: 2023-09

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2024-12-30

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl