

Quality Control Certificate



IRT Neonatal Luminescence Immunoassay

REF RE68019


LOT LIT207

2028-06-30
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value RLU	Range RLU	RLU/RLUmax
CAL A	0	39635	-	3%
CAL B	20	132200	-	10%
CAL C	36	212010	-	15%
CAL D	71	531185	-	38%
CAL E	141	905670	-	65%
CAL F	290	1385245	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	59	43	62	80
CONTROL 2	101	79	112	146



Kit Composition:

REF		LOT	
KLIT655	CONTROL 1	306	2028-06
KLIT665	CONTROL 2	307	2028-06
KLIT511	MTP	K857	2028-06
KLIT545	CONC ENZCONJ	K891	2028-07
KLIT615	CAL A-F	305	2028-06

REF		LOT	
KLIT755	CONC ASSAYBUF	K860	2028-06
KLZZ635	LUMINREAG	K878	2028-06
KLZZ879	CONC WASHBUF	K852	2028-06

Valid edition of the instructions for use: 2024-10

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-22

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl