

Quality Control Certificate



ANA HEp-2 Screen ELISA

REF RE70151	LOT 23305	2026-07-31 YYYY-MM-DD
--------------------	------------------	---------------------------------

Final Release Results: _____

	Result IgG OD	Acceptable range IgG OD / Ratio
CONTROL -	0.024	< 0.200 / < 0.6
CONTROL +	1.553	> 0.800 / 2.1 - 4.5
CUT OFF CONTROL	0.432	> 0.250 / -

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-01-24

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl