

Quality Control Certificate



Rheumatoid Factor IgA/IgG/IgM ELISA

REF	RE70341	LOT	24370		2026-09-30 YYYY-MM-DD
-----	---------	-----	-------	--	--------------------------

Final Release Results:

	Result OD			Acceptable range	
	IgA	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.025	0.021	0.024	< 0.200	-
CAL F	2.180	2.063	2.065	> 1.300	-
CO-CAL	0.460	0.444	0.468	> 0.250	> 0.250

	Result U/mL			Acceptable range	
	IgA	IgG	IgM	U/mL quantitative	Ratio qualitative
CONTROL +	52	52	50	20 - 80	1.2 - 3.3
CONTROL -	1	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2024-11-21

This document has been produced electronically and is valid without a signature.