

Quality Control Certificate



Rheumatoid Factor IgA/IgG/IgM ELISA

REF RE70341

LOT 26060



2028-03
YYYY-MM-DD

Final Release Results:

	Result OD			Acceptable range	
	IgA	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.049	0.033	0.018	< 0.200	-
CAL F	2.313	2.002	2.239	> 1.300	-
CO-CAL	0.438	0.404	0.464	> 0.250	> 0.250

	Result U/mL			Acceptable range	
	IgA	IgG	IgM	U/mL quantitative	Ratio qualitative
CONTROL +	46	47	46	20 - 80	1.2 - 3.3
CONTROL -	1	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-07-16

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl