

# Quality Control Certificate



## Phospholipid IgG/IgM Screen ELISA

<b>REF</b> RE70411	<b>LOT</b> 25110	<b>2027-02-28</b> YYYY-MM-DD
--------------------	------------------	---------------------------------

### Final Release Results:

	Result OD		Acceptable range	
	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.079	0.023	< 0.200	-
CAL F	2.117	2.299	> 1.300	-
CO-CAL	0.432	0.425	> 0.250	> 0.250

	Result U/mL		Acceptable range	
	IgG	IgM	U/mL quantitative	Ratio qualitative
CONTROL +	50	48	20 - 80	1.2 - 3.3
CONTROL -	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-06-24

**Accepted**

This document has been produced electronically and is valid without a signature.