

Quality Control Certificate



Phospholipid IgG/IgM Screen ELISA

| | | |
|--------------------|------------------|---------------------------------|
| REF RE70411 | LOT 25500 | 2027-12-31 YYYY-MM-DD |
|--------------------|------------------|---------------------------------|

Final Release Results:

| | Result OD | | Acceptable range | |
|--------|--------------|-------|------------------|----------------|
| | IgG | IgM | OD quantitative | OD qualitative |
| CAL A | 0.087 | 0.045 | < 0.200 | - |
| CAL F | 2.627 | 2.526 | > 1.300 | - |
| CO-CAL | 0.535 | 0.589 | > 0.250 | > 0.250 |

| | Result U/mL | | Acceptable range | |
|-----------|----------------|-----|-------------------|-------------------|
| | IgG | IgM | U/mL quantitative | Ratio qualitative |
| CONTROL + | 55 | 59 | 20 - 80 | 1.2 - 3.3 |
| CONTROL - | 1 | 1 | < 10 | < 0.6 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-03-13

This document has been produced electronically and is valid without a signature.