

# Quality Control Certificate



## Phospholipid IgG/IgM Screen ELISA

<b>REF</b> RE70411	<b>LOT</b> 26080	<b>2028-02-29</b> YYYY-MM-DD
--------------------	------------------	---------------------------------

### Final Release Results:

	Result OD		Acceptable range	
	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.148	0.073	< 0.200	-
CAL F	2.513	2.455	> 1.300	-
CO-CAL	0.578	0.461	> 0.250	> 0.250

	Result U/mL		Acceptable range	
	IgG	IgM	U/mL quantitative	Ratio qualitative
CONTROL +	46	62	20 - 80	1.2 - 3.3
CONTROL -	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

**Accepted**

Validation date: 2026-04-24

This document has been produced electronically and is valid without a signature.