

Quality Control Certificate



Cholin IgG/IgM ELISA

REF RE70461	LOT 25170	2027-09-30 YYYY-MM-DD
--------------------	------------------	---------------------------------

Final Release Results:

	Result OD		Acceptable range	
	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.050	0.020	< 0.200	-
CAL F	2.123	2.121	> 1.300	-
CO-CAL	0.535	0.433	> 0.250	> 0.250

	Result U/mL		Acceptable range	
	IgG	IgM	U/mL quantitative	Ratio qualitative
CONTROL +	43	45	20 - 80	1.2 - 3.3
CONTROL -	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-03-13

This document has been produced electronically and is valid without a signature.