

Quality Control Certificate



Cardiolipin-Ab IgG / IgM ELISA

REF RE70521	LOT 25450	2027-10-31 YYYY-MM-DD
-------------	-----------	--------------------------

Final Release Results:

	Result OD		Acceptable range	
	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.051	0.035	< 0.200	-
CAL F	2.245	2.186	> 1.300	-
CO-CAL	0.482	0.456	> 0.250	> 0.250

	Result		Acceptable range	
	GPL/mL IgG	MPL/mL IgM	GPL/MPL/mL quantitative	Ratio qualitative
CONTROL +	50	51	20 - 80	1.2 - 3.3
CONTROL -	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-01-09

This document has been produced electronically and is valid without a signature.