

Quality Control Certificate



Cardiolipin-Ab IgG / IgM ELISA

REF	RE70521	LOT	26020		2027-12-31 YYYY-MM-DD
-----	---------	-----	-------	--	--------------------------

Final Release Results:

	Result OD		Acceptable range	
	IgG	IgM	OD quantitative	OD qualitative
CAL A	0.130	0.107	< 0.200	-
CAL F	2.697	2.337	> 1.300	-
CO-CAL	0.515	0.479	> 0.250	> 0.250

	Result		Acceptable range	
	GPL/mL IgG	MPL/mL IgM	GPL/MPL/mL quantitative	Ratio qualitative
CONTROL +	72	62	20 - 80	1.2 - 3.3
CONTROL -	1	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-08-11

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl