Quality Control Certificate

SLA-Ab ELISA

REF	LOT	
RE70791	**25205**	**2028-04-30** YYYY-MM-DD

Final Release Results:

	Result OD	Acceptable range OD quantitative	OD qualitative
CAL A	0.022	< 0.200	-
CAL F	2.475	> 1.300	-
CO-CAL	0.583	> 0.250	> 0.250

	Result U/mL	Acceptable range U/mL quantitative	Ratio qualitative
CONTROL +	50	20 - 80	1.2 - 3.3
CONTROL -	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-10-16

This document has been produced electronically and is valid without a signature.

IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl