

Quality Control Certificate



Intrinsic Factor-Ab ELISA

| | | |
|--------------------|------------------|---------------------------------|
| REF RE70881 | LOT 26080 | 2028-03-31 YYYY-MM-DD |
|--------------------|------------------|---------------------------------|

Final Release Results:

| | Result OD | Acceptable range | |
|--------|--------------|------------------|----------------|
| | | OD quantitative | OD qualitative |
| CAL A | 0.019 | < 0.200 | - |
| CAL F | 2.396 | > 1.300 | - |
| CO-CAL | 0.417 | > 0.250 | > 0.250 |

| | Result U/mL | Acceptable range | |
|-----------|----------------|-------------------|-------------------|
| | | U/mL quantitative | Ratio qualitative |
| CONTROL + | 45 | 20 - 80 | 1.2 - 3.3 |
| CONTROL - | 1 | < 10 | < 0.6 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-04-08

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl