

# Quality Control Certificate



## Parietalcell-Ab ELISA

<b>REF</b> RE70891	<b>LOT</b> 25120	<b>2027-02-28</b> YYYY-MM-DD
--------------------	------------------	---------------------------------

### Final Release Results:

	Result OD	Acceptable range	
		OD quantitative	OD qualitative
CAL A	0.025	< 0.200	-
CAL F	2.590	> 1.300	-
CO-CAL	0.435	> 0.250	> 0.250

	Result U/mL	Acceptable range	
		U/mL quantitative	Ratio qualitative
CONTROL +	53	20 - 80	1.2 - 3.3
CONTROL -	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

**Accepted**

Validation date: 2025-05-15

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)