

Quality Control Certificate



Parietalcell-Ab ELISA

REF RE70891

LOT 25125

2028-02-29
YYYY-MM-DD

Final Release Results:

	Result OD	Acceptable range	
		OD quantitative	OD qualitative
CAL A	0.025	< 0.200	-
CAL F	1.993	> 1.300	-
CO-CAL	0.343	> 0.250	> 0.250

	Result U/mL	Acceptable range	
		U/mL quantitative	Ratio qualitative
CONTROL +	44	20 - 80	1.2 - 3.3
CONTROL -	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl