

Quality Control Certificate



LKM-1-Ab ELISA

REF RE70911	LOT 25060	2027-01-31 YYYY-MM-DD
--------------------	------------------	---------------------------------

Final Release Results:

	Result OD	Acceptable range	
		OD quantitative	OD qualitative
CAL A	0.035	< 0.200	-
CAL F	2.410	> 1.300	-
CO-CAL	0.450	> 0.250	> 0.250

	Result U/mL	Acceptable range	
		U/mL quantitative	Ratio qualitative
CONTROL +	53	20- 80	1.2-3.3
CONTROL -	1	< 10	< 0.6

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-04-22

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl