

Quality Control Certificate



Cardiolipin IgG ELISA

REF RE75011

LOT 26060058

2027-06-30
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	GPL-U/mL	OD	lower limit OD	target value OD	upper limit OD
CAL A	0	0.064			< 0.100
CAL F	100	1.992	> 1.000		
CONTROL -	-	0.099			≤ 0.300
CONTROL +	-	0.858	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value GPL-U/mL	lower limit GPL-U/mL	target value GPL-U/mL	upper limit GPL-U/mL
CONTROL -	1.2			≤ 6.0
CONTROL +	31	19	30	41

Factor for qualitative assay evaluation	0.48
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-06

Accepted

This document has been produced electronically and is valid without a signature.