


Quality Control Certificate



Cardiolipin IgM ELISA

REF RE75021

LOT 25110156

 2027-02-28
YYYY-MM-DD

Final Release Results:

	Conc. MPL-U/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
CAL A	0	0.053			< 0.100
CAL F	100	1.956	> 1.000		
CONTROL -	-	0.103			≤ 0.300
CONTROL +	-	1.044	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value MPL-U/mL	Acceptable range		
		lower limit MPL-U/mL	target value MPL-U/mL	upper limit MPL-U/mL
CONTROL -	1.1			≤ 5.0
CONTROL +	24	18	28	38

Factor for qualitative assay evaluation	0.31
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-01-23

This document has been produced electronically and is valid without a signature.