

Quality Control Certificate



Cardiolipin IgM ELISA

REF RE75021

LOT 26050133

 2027-06-30
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	MPL-U/mL	OD	lower limit OD	target value OD	upper limit OD
CAL A	0	0.058			< 0.100
CAL F	100	2.082	> 1.000		
CONTROL -	-	0.105			≤ 0.300
CONTROL +	-	1.306	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value MPL-U/mL	lower limit MPL-U/mL	target value MPL-U/mL	upper limit MPL-U/mL
CONTROL -	0.9			≤ 5.0
CONTROL +	31	18	28	38

Factor for qualitative assay evaluation	0.31
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-06

Accepted

This document has been produced electronically and is valid without a signature.