Quality Control Certificate

TECAN.

beta2-Glycoprotein 1 IgG ELISA

REF	LOT	
RE75051	25090226	2026-09-30 YYYY-MM-DD

Final Release Results:

	Conc. U/mL	Measured value OD	Acceptable range lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.056			< 0.100
Calibrator F	100	1.888	> 1.000		
CONTROL -	-	0.093			≤ 0.300
CONTROL +	-	0.989	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value U/mL	Acceptable range lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	1.6			≤ 6.0
CONTROL +	40	26	40	54

Factor for qualitative assay evaluation	0.37

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-10-10

This document has been produced electronically and is valid without a signature.

IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl