


# Quality Control Certificate



## beta2-Glycoprotein 1 IgM ELISA

REF RE75061

LOT 25110065

 2026-10-31  
YYYY-MM-DD

### Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
Calibrator A	0	0.057			< 0.100
Calibrator F	100	2.297	> 1.000		
CONTROL -	-	0.117			≤ 0.300
CONTROL +	-	1.263	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
		lower limit	target value	upper limit
U/mL	U/mL	U/mL	U/mL	
CONTROL -	1.5			≤ 6.0
CONTROL +	36	22	35	48

Factor for qualitative assay evaluation	0.45
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

# Accepted

Validation date: 2025-12-08

This document has been produced electronically and is valid without a signature.