

Quality Control Certificate



beta2-Glycoprotein 1 IgM ELISA

REF RE75061

LOT 26070098

2027-06-30
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL	OD	lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.060	> 1.000		
Calibrator F	100	2.590			
CONTROL -	-	0.122	≥ 0.500		
CONTROL +	-	1.376			

	Validation criteria (manual)			
	Measured value	Acceptable range		
	U/mL	lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	1.4	22	35	≤ 6.0
CONTROL +	35			48

Factor for qualitative assay evaluation	0.45
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-16

Accepted

This document has been produced electronically and is valid without a signature.